



Medical Silicone Keyboard Care and Cleaning Protocol

Compatible Chemicals:

- OPA solution (0.55%)
- Household bleach (5%) 5,000 ppm
- Alcohol (up to 85%)
- Hydrogen peroxide
- Accelerated hydrogen peroxide (.01% – 7.0%)
- Quaternary Ammonia Compounds
- Most detergents



Hard Backed Model #79700BK

Includes on/off function for easy cleaning

To Activate: press the on/off button while pressing the green function key

Recommended:

For Cleaning:

- Most detergents
- Stabilized hydrogen peroxide (.02%)
- Accelerated hydrogen peroxide (.02% - .05%)



Flexible Model #79500

For Sanitizing (proven bacterial efficacy):

- Accelerated hydrogen peroxide (.02% - .05%)
- Alcohols (+60% isopropyl or ethyl alcohol)
- Bleach (1% bleach, 99% water)

For Low Level Disinfecting (terminal):

- Accelerated hydrogen peroxide (0.05%) in 5 minutes
- Quaternary ammonia compounds with alcohol (10% - 20%) in 5 minutes
- Bleach (500-1000 ppm)

****NOW AVAILABLE- Optical Silicone Mouse!!**



Do Not Use:

- Petroleum based cleaning fluids such as oil, gasoline, turpentine
- Acetone, toluol or other corrosives
- Cleaning fluids containing more than 10% bleach
- Temperature sterilization
- Abrasive scrubbing pads or buffers

TO AVOID MIXING CHEMICALS, WE RECOMMEND TRYING TO SELECT PRODUCTS THAT ARE BOTH GOOD CLEANERS AND GOOD DISINFECTANTS

For use in temperatures of -40F to +80F

Do not fold, bend, twist or pull the keyboard or place heavy objects on it. Avoid sharp objects.

CAESA - LAB INC.

4046 Rue de l'orchidée, Saint-Bruno, Quebec, J3V 6J9, CANADA
Tel: 450-441-5408 Fax: 450-441-9167